



EKOTEKS

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE



AB-0583-T

2400009433-
RER

03.25

TEST REPORT
DENEY RAPORU

Report Nr / Revision Nr : 2400009433
Customer name : ALDO BEBEK ÜRÜNLERİ SANAYİ VE TİCARET LTD.ŞTİ
Applicant Address : AKTİM AKÇAB URGAZ MH AKÇABURGAZ CD
/ İSTANBUL
Buyer name : -
Contact Person : -
Sample Accepted on : 22.11.2024
Re-submitted/re-confirmation date : -
Date of test : 22.11.2024 / 19.12.2024
Fiber Composition : Claimed to be :-
Sample Description : WHITE BABY CLEANING COTTON + BREAST PAD
Order No : -
Model Number : 926-Baby cleaning cotton
1021-Breast Pad 30+10'lu
Previous Report No : -
Sampling : The results given in this report belong to the received sample by vendor.
End use : -
Care label : -
Decision Rule : -
Disclaimer Statement : -
Conformity Assessment : -

The Turkish Accreditation Agency (TÜRKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports. EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.

The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.

Seal

Date
06.03.2025

Customer Representative
YAGMUR OKUMUS

Laboratory Manager
Sevim A. RAZAK
06.03.2025



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T
2400009433 -RER
03.25

TEST RESULTS

TEST	METHOD	EVALUATION
ABSORPTION FOR NON-WOVEN SURFACES (*)	ISO 9073-6:2000(*)	-
BIOTECHNOLOGY	ISO 11737	P

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified.If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The declaration of conformity was given in accordance with the Simple Acceptance Decision Rule. (without considering the level of confidence and measurement uncertainty , evaluation of suitability or non-conformity based on whether the test result obtained is only within the specified limits) Tests marked (*) in this report are not included in the accreditation schedule.

Note: The report dated 17.12.2024 dated 2400009433 has been added to the 2400009433 -RER due to the change of "Customer name" and has been replaced with the report dated 06.03.2025

TEST RESULTS

Photo of the sample received



TEST RESULTS

Photo of the sample received



TEST RESULTS

Photo of the sample received



EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş.

AB-0583-T


2400009433

-RER

03.25

TEST RESULTS

Photo of the sample received



TEST RESULTS

Photo of the sample received



TEST RESULTS

Photo of the sample received



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

2400009433
-RER

03.25

TEST RESULTS

Photo of the sample received



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T
2400009433 -RER
03.25

TEST RESULTS

Sample ID	Part No.	Part ID	Explanatory Information
A	1	Göğüs Pedi	

Summary of Results

Parts & Results															
Parts	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
BIOTECHNOLOGY	P	-	-	-	-	-	-	-	-	-	-	-	-	-	-

- P Pass
F Fail
LS Lack of sample.
NA Not applicable
R Refer to retailer technologist

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T
2400009433 -RER
03.25

TEST RESULTS

SPECIMEN DESCRIPTION :

ABSORPTION FOR NON-WOVEN SURFACES

ISO 9073-6:2000(*)

Performed in the conditioned room (20±2°C-65±4 % RH)

Grade 3 type water was used in test.

Parameters	RESULT	REQUIREMENT	EVALUATION
Part 1 (Liquid Absorbency Time) (s):	2	-	-
Dimensions of the test sample Part 1 : Width : 76 , Length : 350 mm			

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T
2400009433 -RER
03.25

TEST RESULTS

Test Details	
Number of specimens	5 pieces
Shaking time and rotation speed	5 min at 250 rpm
Total viable aerobic microbial count (PCA)	Temperature: 30±1°C Time: 72 hour PCA medium (Lot no : 0923)
Total fungi count (SDA)	Temperature: 20°C Time: 5 day SDA medium (Lot no : 0923)
Total anaerobic microbial count (Kanlı agar)	Temperature: 30±1°C Time: 72 hour Blood Agar medium (Lot no : 0923)
Used Test Solution	Sodium chloride, peptone buffer

RESULT							REQUIREMENT
Number of Specimen	Weight (g)	Aerobic cfu/100 ml	Fungal cfu/100ml	Anaerobic cfu/100ml	Total Bioburden cfu/sample	Total Bioburden cfu/g	Total Microbial Load
1	10x10	9	3	12	72	0,72	≤30 cfu/g
2	10x10	6	1	7	42	0,42	
3	10x10	8	2	10	60	0,6	
4	10x10	5	1	6	36	0,36	
5	10x10	7	2	9	54	0,54	

*kob: Colony forming unit

Biological Burden Correction Factor:
Total Uncertainty: 0,10 (U)